

Medical Device Reporting

**FDA Small Business
Regulatory Education for Industry (REdI)
Burlingame, CA
May 16, 2018**

Anike Freeman

Senior Consumer Safety Officer (Acting)
Division of Industry and Consumer Education
Office of Communication and Education
Center for Devices and Radiological Health
U.S. Food and Drug Administration






Learning Objectives

- Establish the intent of 21 CFR 803
- Identify basic reporting requirements
- Discuss harmonized adverse event codes
- Understand how FDA utilizes MDRs

MDRs Stem from Complaint Files



Value of Medical Device Reports

Consumers and Industry:

- Device safety and performance
- Design improvement

FDA:

- Monitor device safety and performance
- Assess need for regulatory action



Medical Device Reporting (MDR)

21 CFR 803

- Supplements provisions of 21 CFR 820
- Establishes requirements for firm's medical device reporting system
 - Standardized complaint review process
 - Timely, effective identification and communication of adverse events
 - Documentation and recordkeeping

21 CFR [803.17](#)

Mandatory Reporting

Reportable Events reasonably suggest a marketed device:

- May have caused or contributed to a death or serious injury, or
- Malfunction was likely to cause or contribute to death or serious injury were it to recur

21 CFR [803.1](#), 21 CFR [803.3](#)

Who Reports?

Mandatory Reporters:

- Manufacturers
- Importers
- Device User Facilities

Voluntary Reporters:

- Patients
- Healthcare Professionals
- Consumers



How are MDRs Submitted?

Mandatory Reporters (Manufacturers):

- Electronic submission only
 - eMDR Final Rule issued August 14, 2015
 - Use Electronic Submissions Gateway (ESG)

Voluntary Reporters:

- Online via MedWatch (Form FDA 3500)
- Mobile app
- Download and submit via postal mail

Adverse Event Codes: Harmonization with IMDRF*

***IMDRF = International Medical Device Regulators Forum**

Adverse Event Coding Working Group

FDA is participating in IMDRF working group

- IMDRF Adverse Event Terminologies
- Developing coding schemes
 - Device Problem Codes
 - Device Evaluation Codes
 - Patient Problem Codes
 - Component Codes

FDA Harmonization Efforts

- Released two code lists in November 2017
 - Device Problem and Manufacturer Evaluation Codes
 - Retires FDA terms not included in the IMDRF lists
 - Adds IMDRF terms not previously in FDA lists
 - New FDA and NCIt codes assigned to new terms

FDA Harmonization Efforts

- New code lists go live July 5th, 2018
- Retired codes will no longer be accepted
- Plan accordingly

Future Coding Efforts

- Currently under development by IMDRF
 - Patient Problem Codes and Component Codes
 - FDA plans to harmonize with those coding sets
- Once harmonization is complete
 - Update internal systems to accept IMDRF alphanumeric codes
 - Continue to accept FDA and NCI codes



When to Report

| Reporter: | What to Report: | Reports to: | When: |
|----------------------|---|-----------------------------|---------|
| Manufacturers | Deaths, serious injuries, or certain malfunctions | FDA (3500A) | 30 days |
| | Events requiring remedial action* | FDA (3500A) | 5 days |
| Importers | Deaths or serious injuries | FDA (3500A) Manufacturer | 30 days |
| | Certain malfunctions | Manufacturer | 30 Days |
| User Facility | Deaths or serious injuries | FDA (3500A) | 10 days |



When to Report

| Reporter: | What to Report: | Reports to: | When: |
|----------------------|---|-----------------------------|---------|
| Manufacturers | Deaths, serious injuries, or certain malfunctions | FDA (3500A) | 30 days |
| | Events requiring remedial action* | FDA (3500A) | 5 days |
| Importers | Deaths or serious injuries | FDA (3500A) Manufacturer | 30 days |
| | Certain malfunctions | Manufacturer | 30 Days |
| User Facility | Deaths or serious injuries | FDA (3500A) | 10 days |

When to Report

| Reporter: | What to Report: | Reports to: | When: |
|----------------------|---|-----------------------------|---------|
| Manufacturers | Deaths, serious injuries, or certain malfunctions | FDA (3500A) | 30 days |
| | Events requiring remedial action* | FDA (3500A) | 5 days |
| Importers | Deaths or serious injuries | FDA (3500A) Manufacturer | 30 days |
| | Certain malfunctions | Manufacturer | 30 Days |
| User Facility | Deaths or serious injuries | FDA (3500A) | 10 days |



Poll:

What happens to submitted MDRs?

- A. Reviewed, analyzed, and redacted by FDA
- B. Entered in MAUDE database
- C. A and B
- D. Nothing, ESG connects directly to a black hole

Statistics

- FDA receives ~855,000 reports per year
 - 936,508 reports in 2017
 - Majority from Mandatory Reporters
 - 7,200 from Voluntary Reporters
- Each report is reviewed by an analyst
- Analysts are medical and technical professionals
 - Nurses, engineers, scientists etc.

Analyst Medical Subject Areas

Product Evaluation Branch I

- Cardiovascular
- Gastroenterology
- Renal/Urology
- OB/GYN

Product Evaluation Branch II

- General Hospital
- General/Plastic Surgery
- Orthopedics
- Infection Control
- Physical Medicine

Product Evaluation Branch III

- Respiratory
- Anesthesia
- Neurology
- Ear/Nose/Throat
- Dental
- Ophthalmology

What are Analysts looking for?

Trends:

- Common vs Novel
- Frequency
- Severity
- Risk



Information Analysis Branch

- CDRH's adverse event data management team
- Maintains MAUDE database and data files
- Provides eMDR Support
 - Houses eMDR Help Desk (eMDR@fda.hhs.gov)

Possible Actions

- Contact the reporter
- Escalate the signal within CDRH
- Make recommendations
 - FDA inspection of manufacturer
 - Changes to device labeling
 - Public health notifications
 - Device recall

Other Uses of MDR Data

- Educate premarket review staff
- Refining premarket review
- Response to congressional inquiries

Summary

- MDRs are a regulatory requirement per 21 CFR 803
- You must understand your reporting obligations
- You must comply with reporting obligations
- MDRs are critical to public health and safety

Resource Websites

- [Medical Device Reporting \(MDR\)](#)
- [Guidance Document: Medical Device Reporting for Manufacturers](#)
- [eMDR Final Rule](#)
- [How to Enroll in eMDR Program](#)
- [Setting up a Web Trader Account Checklist](#)
- [eSubmitter Download and Installation](#)
- [Health Level Seven \(HL7\) Individual Case Safety Reporting](#)
- [CDRH Learn](#)

Resource Websites

- [MDR Adverse Event Codes](#)
- [How to Code](#)
- [Coding Resources](#)
- [Coding FAQs](#)
- [eMDR Help and FAQs](#)

Resources

- **General MDR questions: Division of Industry and Consumer Education (DICE)**

Email: DICE@fda.hhs.gov

Phone: (800) 638-2041

- **Interpretations of MDR policy: MDR Policy Group**

Phone: (301) 796-6670 (voice)

Email: MDRPolicy@fda.hhs.gov

Questions

Please evaluate this session:

surveymonkey.com/r/DEV-D2S07

Your Call to Action

- Understand and comply with 21 CFR 803
- Utilize harmonized adverse event codes
- Remember your reports support public health

