

510(k) Submission Types and Reasons for Conversion

FDA Small Business Regulatory Education for Industry (REdI)

June 8, 2022

Melissa Hall

Policy Analyst

Office of Regulatory Programs

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

U.S. Food and Drug Administration

A man with dark hair, wearing a white shirt, is shown from the side, resting his head on his hand. He is holding a white rectangular sign with the word "help" written in a large, black, handwritten-style font. The background is a plain, light-colored wall.

help

Learning Objectives

- Provide an overview of the 510(k) process
- Describe the 510(k) Submission Types
- Discuss some common reasons for conversion
- Apply concepts to real-life example

Overview of 510(k) Process

What is a 510(k)?



A 510(k) is a **premarket submission** made to FDA to demonstrate that the device to be marketed is at least as safe and effective, that is, **substantially equivalent, to a legally marketed device** (21 CFR 807.92(a)(3)) that is not subject to PMA*.

Fun Fact:

It is the biggest premarket program in CDRH

- FDA receives ~3000 510(k) submissions per year
- ~90% are found SE and go to market

*21 CFR 807.92(a)(3)

When is a 510(k) Required?


- Introducing device to market for the first time
- Modification(s) to device that could affect safety or effectiveness:
 - Intended use
 - labeling
- Reprocessing of a single use device (SUD)*

2017 Guidance: “Deciding When to Submit a 510(k) for Change to an Existing Device”

***FDA Guidance: Frequently-Asked-Questions about the Reprocessing and Reuse of Single-use devices by Third-party and Hospital Reprocessors**

High-Level Process Overview

510(k) Submission Core Process



Acronyms:

RTA = Refuse to Accept

LR = Lead Reviewer


SI = Substantive Interaction

PI = Proceed Interactively

AI = Additional Information

FDA Guidance: Refuse to Accept Policy for 510(k)s

The 510(k) Flowchart



- 2014 Guidance: Evaluating Substantial Equivalence in Premarket Notifications
- Flowchart not intended to be used as a stand-alone document
- Decision questions are answered in order
- SE = answer all decision questions with a predicate.

Describe 510(k) Submission Types

510(k) Submission Types

- Traditional
- Abbreviated
 - Safety & Performance (S&P) Based Pathway
- Special
- Third Party

510(k) Submission Types

Commonalities

- All 510(k) submissions must identify a predicate
- All 510(k) submissions include content to demonstrate SE to the cited predicate(s)

Differences

- FDA review timeframe goals
- Different RTA criterion and checklists for different submission types
- The nature of the content can vary depending on 510(k) submission type
 - Summary reports vs. full reports

Traditional 510(k)

- Most common
- Used for any original 510(k) or for a change to a previously cleared device under 510(k)
- **All** data provided:
 - full test reports, etc.
- 90-day FDA review clock

FDA Guidance: Format for Traditional and Abbreviated 510(k)s

Abbreviated 510(k)

- Device relies on:
 - FDA guidance document(s)
 - Demonstration of compliance with special controls for the device type
 - Voluntary consensus standard(s)
 - Guidance- [“Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices”](#)
- Summary reports
- 90-day FDA review clock

Safety and Performance Based Pathway:

Overview

- Expands on existing Abbreviated 510(k) Program
- Removes requirement for direct predicate comparison testing for some performance characteristics
 - Meet FDA-identified performance criteria to demonstrate device is as safe and effective as predicate device
- 90-day FDA review clock

FDA Guidance: Safety and Performance Based Pathway

Safety and Performance Based Pathway:

Eligible Device Types

- Currently there are 7 device types appropriate for this pathway:
 - Spinal Plating System
 - Orthopedic Non-spinal Metallic Bone Screws and Washers
 - Cutaneous Electrodes for Recording Purposes
 - Conventional Foley Catheters
 - Fracture Fixation Plates
 - Surgical Sutures
 - Denture Base Resin-NEW!
 - Facet Screw Systems-NEW!
- Draft guidance:
 - Soft (Hydrophilic) Daily Wear Contact Lenses

Docket Number FDA-2018-D-1387 at www.regulations.gov

Safety and Performance Based Pathway:

Eligibility Criteria

- The new device has
 - Same **indications for use**
 - Same **technological characteristics**
 - Does not raise different questions of safety and effectiveness
- Meets all the **FDA-identified performance criteria**

Special 510(k) Program

- Change to your device (IFU/labeling or technology)
- Performance data:
 - Not needed
 or
 - Well-established methods exist to evaluate the change
 - Can be reviewed in summary or risk analysis format
- 30-day FDA review clock

Well-Established Methods

- Previously-cleared 510(k)
- Recognized consensus standard
- Widely available/accepted methods/premarket submission



Summary/Risk Analysis

| Device Change | Risks | V & V Method | Acceptance Criteria | Summary of Results |
|---------------|-------|--------------|---------------------|--------------------|
| | | | | |

- Summary of design control procedures (21 CFR 820.30)
 - Results from verification and validation (V&V) activities
- SE determination can be made with summary information
- Examples outlined in Appendix C of guidance

Third Party Review Program

- Traditional, Abbreviated, or Special
- Review completed by accredited 3rd party*
- Certain low/moderate risk devices are eligible*
- 30 FDA days for supervisory review
- 3P510k@fda.hhs.gov

| | Traditional | Abbreviated | S&P Pathway | Special | Third Party |
|--------------------------------------|-------------------|---|---|---|--|
| Overview of differences/similarities | Most common | Similar to Traditional in terms of timeline | Expansion of Abbreviated | Used for modifications to own device | Review of certain low/moderate risk devices. FDA performs supervisory review |
| FDA review time | 90-days | 90-days | 90-days | 30-days | 30-days |
| Data needs | All data provided | Relies on standards, guidance, special controls | Meets FDA identified performance criteria | Only Summary -level information focusing on the modifications | Relies on standards, guidance, special controls |
| Guidance | yes | yes | yes | yes | yes |

Knowledge Check

What is the FDA MDUFA review time for an S&P Pathway submission?

- a. 30-day
- b. 60-day
- c. 90-day

Common Reasons for Conversion

Conversion: Special to Traditional



- No well-established methods
- Summary or Risk Analysis format is not appropriate for the proposed change(s).
- The subject device is not the submitters own device

Conversion: Abbreviated to Traditional

- Summary reports provided that do not reference:
 - Guidance
 - Special controls
 - Standards

Conversion Case Study

Scenario

| | Cleared Device | Current Device |
|----------------------|---|---|
| |  |  |
| Indications for Use | Same | Same |
| Material | Same | Same |
| Sterilization Method | Non-sterile | Sterilized via gamma irradiation |
| Screw length | 30-60mm | 25-60mm |

Special 510(k)




Figure 1. Special 510(k) flowchart.

A - Is it a change to the manufacturer's own device?

B – Are performance data needed to evaluate the change?

C – Is there a well-established method to evaluate the change?

D – Can the data be reviewed in a summary or risk analysis format?

Performance Data Needed?

| Device Change | Risks | V & V Method | Acceptance Criteria | Summary of Results |
|----------------------------------|--------------------------------------|--|--|---|
| Sterilized via gamma irradiation | Potential for infection or rejection | biocompatibility, sterility, pyrogenicity, package integrity, and shelf-life | FDA guidance | Scientifically-based rationale supporting no further biocomp testing needed |
| Smaller Screw length | Potential to create new worst-case | Finite Element Analysis/Mechanical testing | FDA guidance/compare testing to FDA cleared device | Summary of results shows no new worst-case |

Well-Established Methods?

| Device Change | Well-established methods |
|----------------------------------|---|
| Sterilized via gamma irradiation | FDA guidance Submission and Review of Sterility Information in Premarket Notification (510(k)) Submissions for Devices Labeled as Sterile |
| Smaller Screw length | FDA guidance: Spinal Systems 510(k)s |

Summary/Risk Analysis Format?

| Device Change | Can the change be evaluated in a summary or risk analysis format? |
|----------------------------------|--|
| Sterilized via gamma irradiation | <p>Yes</p> <ul style="list-style-type: none"> • The methods are standardized, and the results can be summarized • SE determination <u>NOT</u> dependent on underlying data, such as images, raw graphs, or line item data |
| Smaller Screw length | <p>No</p> <ul style="list-style-type: none"> • Methods outlined in FDA guidance • Results CANNOT be summarized • SE determination dependent on underlying data such as images, raw graphs, or line item data |

Special 510(k)




Figure 1. Special 510(k) flowchart.

No, there is a change that is inappropriate for review as a special 510(k).

Knowledge Check

If the device change was ONLY the sterilization method, could it have been reviewed as a Special?

- a. Yes
- b. No
- c. It depends

Resources



| Slide # | Cited Resource | URL |
|---------|--|--|
| 5 | 21 CFR 807.92(a)(3) | www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfCFR/CFRSearch.cfm?r=807.92 |
| 6 | Deciding When to Submit a 510(k) for Change to an Existing Device | www.fda.gov/regulatory-information/search-fda-guidance-documents/deciding-when-submit-510k-change-existing-device |
| 6 | Frequently-Asked-Questions about the Reprocessing and Reuse of Single-Use Devices by Third-Party and Hospital Reprocessors | www.fda.gov/regulatory-information/search-fda-guidance-documents/frequently-asked-questions-about-reprocessing-and-reuse-single-use-devices-third-party-and-hospital |
| 7 | Refuse to Accept Policy for 510(k)s | www.fda.gov/regulatory-information/search-fda-guidance-documents/refuse-accept-policy-510ks |
| 12 | Format for Traditional and Abbreviated 510(k)s | www.fda.gov/regulatory-information/search-fda-guidance-documents/format-traditional-and-abbreviated-510ks |
| 13 | The Abbreviated 510(k) Program guidance | www.fda.gov/regulatory-information/search-fda-guidance-documents/abbreviated-510k-program |

Resources



| Slide # | Cited Resource | URL |
|---------|--|--|
| 16 | The Special 510(k) Program | www.fda.gov/regulatory-information/search-fda-guidance-documents/special-510k-program |
| 18 | Quality System (QS) Regulation/Medical Device Good Manufacturing Practices | www.fda.gov/medical-devices/postmarket-requirements-devices/quality-system-qs-regulationmedical-device-good-manufacturing-practices |
| 19 | List of Devices for Third Party Review | www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfThirdParty/current.cfm |
| 19 | Current List of FDA-Recognized 510(k) Third Party Review Organizations | www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfthirdparty/accredit.cfm |
| 19 | 510(k) Third Party Review Program | www.fda.gov/regulatory-information/search-fda-guidance-documents/510k-third-party-review-program |

Summary

- Provided a high-level overview of 510(k) process
- Described 510(k) Submission Types and Third Party review program
- Discussed some common reasons for conversion
- Applied key concepts to a real-life conversion example

Questions

